

Queensland Paediatric ETT Size Guide

Clinicians can take the age and size of the child into account when selecting a suitable ETT for intubation and move up or down in size as required during the procedure.

Age	ETT micro-cuff size I.D. (mm) ± 0.5	ETT size I.D. (mm) ± 0.5	Oral length (cm)	LMA	Laryngoscope blade	NGT size (Fg)
0-6 months	3	3	9.0-10.5	1	0	6-8
6-12 months	3.5	3.5	10.5-12.5	1.5	0	8
1 year	3.5	4	12.5	1.5	1	8
2 years	4	4.5	13	2	2	10
3 years	4	4.5	13.5	2	2	10
4 years	4.5	5	14	2.5	2	10
5 years	4.5	5	14.5	2.5	2	10
6 years	5	5.5	15	2.5	2	10-12
7 years	5	5.5	15.5	2.5	2	10-12
8 years	5.5	6	16	3	2	12
9 years	5.5	6	16.5	3	2	12
10 years	6	6.5	17	3	3	12
11 years	6	6.5	17.5	3	3	12
12 years	6.5	7	18	4	3	12
13 years	6.5	7	18.5	4	3	12-14
14 years	7	7.5	19	4	4	12-14
15 years	7	7.5	19.5	4	4	14
16 years	8	8	20	4	4	14

